

Laboratoire Huckert's International srl

Avenue Lavoisier 20, 1300 Wavre, Belgium

has been assessed and certified as meeting the requirements of

Directive 93/42/EEC on medical devices, Annex II (excluding Section 4)

For the following products

The scope of registration appears on page 2 of this certificate

This certificate is valid from 21 May 2021 until 25 July 2023
and remains valid subject to satisfactory surveillance audits.

Issue 4. Certified since 25 July 1999

Certification is based on reports numbered BE/AND 200467

Authorised by



Global Medical Devices Head of Notified Body

SGS Belgium NV, Notified Body 1639

SGS House Noorderlaan 87 2030 Antwerp Belgium

t +32 (0)3 545 48 48 f +32 (0)3 545 48 49 www.sgs.com

LPMD5007 - Certificate CE1639 Annex II-4 - EN rev. 02

Page 1 of 2



Laboratoire Huckert's International srl

Directive 93/42/EEC

on medical devices, Annex II (excluding Section 4)

Issue 4

Detailed scope

**Concentrated UMONIUM38® dedicated to disinfection
of invasive medical devices (Surgical instruments)
and surface disinfection of non-invasive medical devices.**

**Ready to be used UMONIUM38® dedicated to disinfection
of invasive medical devices (Surgical instruments)
and surface disinfection of non-invasive medical devices.**

Details

Surface disinfectants for non invasive medical devices

- Umonium 38 ® Equipments (125ml, 1L, 5L, 25L)
- Umonium 38 ® Medical Tissues (1, 10, 100, 95)
- Umonium 38 ® Neutralis tissue (1, 100, 95 wipes)

**Disinfectants for invasive and non-invasive medical devices,
with the exclusion of contact lenses**

- U38 Instrument (125ml, 1L and 5L)
- U38 Instrument & Equipment (125ml, 1L and 5L)
- U38 Medical Spray (spray 250ml, 500ml and 1L)
- U38 Labocid (25L) - U38 Neutralis (125ml, 1L and 5L)
- U38 Neutralis Spray (250 and 500 ml)
- U38 Sterily (125ml, 1L and 5L)

Where the above scope includes class III medical device(s), a valid EC Design Examination Certificate according to Annex II (Section 4) is a mandatory requirement for each device in addition to this certificate to place that device on the market.

Laboratoire Huckert's International srl
20, Avenue Lavoisier
1300 Wavre
Belgium

18 Jul 2023

Confirmation Letter Reference: CLNB1639 – BE/AND/200467.QMD

To whom it may concern,

Confirmation of receipt of a formal application and conclusion of written agreement in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

This letter confirms that, SGS Belgium NV, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number 1639 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Laboratoire Huckert's International srl
20, Avenue Lavoisier
1300 Wavre
Belgium
SRN Number: BE-MF-00024078

The devices covered by the formal application and the written agreement mentioned above are shown at the end of this letter.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired prior to the publication of the Regulation EU 2023/607 on 15th March 2023, this letter also confirms that:

- The manufacturer submitted the MDR application and signed the written agreement by the date of MDD/AIMDD certificate expiry;
- The certificates expired after 26th May 2021 by course of time and were valid at the date of their expiry neither having been suspended nor withdrawn.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3 of MDR (as amended by EU 2023/607), are shown below:

- 26th May 2026 for Class III custom-made implantable devices
- 31st December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31st December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31st December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body SGS Belgium NV 1639,



pp [Jérôme JADOT]

Virginie SILORET
Global Medical Device Certification Manager
Email: Virginie.siloret@sgs.com
Phone: +41 22 739 98 58

Devices covered by this letter: